AUTOCLAVE RECORD
Non-Media

Year: _____

Date	Contents	Time In	Time at 121±1°C	Time Out	Total Time in autoclave	MRT/D.L. Temp °C	Initials	Remarks

Membrane filters and pads	10 min	Dilution water blanks	15 min
Membrane filter assemblies	15 min	Rinse water 500-1000mL	45 min
Sample collection bottles	15 min	Rinse water >1000mL	**
Individual glassware	15 min	Contaminated waste	30 min

** Time adjusted for volume as documented in Quality Assurance Manual

D.L. =Data logger