AUTOCLAVE QUALITY CONTROL
Monthly/Quarterly

Spore Manufacturer: _____ Year: _____
Expiration Date: _____
Lot Number: _____

Note: Check autoclave timer quarterly

Date	Contents	MRT/ D.L. °C	Spore Control (G/ NG)		Spore Test (G/ NG)		Comments	Initials

Date	Stopwatch	Timer	Comments	Initials

G =Growth, NG = No Growth

Incubation Time: _____ Incubation Temperature: _____°C

Acceptance Criteria:
Spore control = Growth
Spore test = No Growth
Timer: ± 1 minute for a 15-minute cycle
D.L.= Data logger